



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-201 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202205-021  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.3           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2022892-201A SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9203018-01N  
Date Received: 02/28/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 03/02/92  
Analysis Date: 03/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.2           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2022892-201B SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9203018-02N  
Date Received: 02/28/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 03/02/92  
Analysis Date: 03/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.3           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2022892-201C SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9203018-03N  
Date Received: 02/28/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 03/02/92  
Analysis Date: 03/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.4           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2022892-201D SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9203017-03N  
Date Received: 02/28/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 03/02/92  
Analysis Date: 03/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.2           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-202 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-08N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021092-203 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202123-01N  
Date Received: 02/11/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/11/92  
Analysis Date: 02/11/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021092-204 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202123-02N  
Date Received: 02/11/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/11/92  
Analysis Date: 02/11/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-205 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-11N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.1           | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-206 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-07N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92

Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.2           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-207 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-10N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-208 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-09N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-209 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-08N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-210 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-06N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-211 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-07M  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.1           | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-212 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-05N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-213 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-06N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021092-214 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202123-031  
Date Received: 02/11/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/11/92  
Analysis Date: 02/11/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.2           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021092-215 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202123-04  
Date Received: 02/11/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/11/92  
Analysis Date: 02/11/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021092-216 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202123-051  
Date Received: 02/11/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/11/92  
Analysis Date: 02/11/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.1           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-217 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-05N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021092-218 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202123-061  
Date Received: 02/11/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/11/92

Analysis Date: 02/11/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-219 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-04N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92

Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-220 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-10N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-221 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-03N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.1           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-222 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-09N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.1           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-223 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-02N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-224 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202202-01N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-225 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202203-11N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-226 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-04N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.2           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-227 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202203-10N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-228 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-03N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021092-229 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202123-07N  
Date Received: 02/11/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/11/92  
Analysis Date: 02/11/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021092-230 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202123-08  
Date Received: 02/11/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/11/92  
Analysis Date: 02/11/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.1           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-231 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-02N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.6           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2022892-231A SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9203018-04N  
Date Received: 02/28/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 03/02/92  
Analysis Date: 03/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2022892-231B SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9203018-05N  
Date Received: 02/28/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 03/02/92  
Analysis Date: 03/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.1           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2022892-231C SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9203018-06N  
Date Received: 02/28/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 03/02/92  
Analysis Date: 03/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2022892-231D SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9203017-07N  
Date Received: 02/28/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 03/02/92  
Analysis Date: 03/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.4           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 203C1692, SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9201072-03N  
Date Received: 01/07/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 01/07/92  
Analysis Date: 01/07/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 1.0 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 202A1792, SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9201074-03N  
Date Received: 01/08/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 01/08/92  
Analysis Date: 01/08/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.2           | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.

Report: P C B P R E

Initials: KF 7/29

Status: D R A F T

Printed: APR. 28 1992 7:45 PM

Client: Clean Harbors of Kingston, Inc.

CHAS Lab #: 9204345-01N

Sample I.D.: 2042192-SRN06 SCUDDER

Date Received: 04/22/92

Sample Type: Wipe

## Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/23/92

Analysis Date: 04/24/92

| Parameter          | MDL | Concentration          | Units        |
|--------------------|-----|------------------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND                     | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND                     | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND                     | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND                     | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND                     | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND                     | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 1.4 - ok <del>KF</del> | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)

Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.

## ----- NOTES -----

Calculated and reported by KW.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-SRNO6A SCUDDER POST  
Sample Type: Wipe

CHAS Lab #: 9205046-01N  
Date Received: 05/06/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/06/92  
Analysis Date: 05/06/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-SRNO6B SCUDDER POST  
Sample Type: Wipe

CHAS Lab #: 9205046-02N  
Date Received: 05/06/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/06/92

Analysis Date: 05/06/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-SRNO6C SCUDDER POST  
Sample Type: Wipe

CHAS Lab #: 9205046-03N  
Date Received: 05/06/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/06/92  
Analysis Date: 05/06/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-SRNO6D SCUDDER POST  
Sample Type: Wipe

CHAS Lab #: 9205046-04N  
Date Received: 05/06/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/06/92  
Analysis Date: 05/06/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-REC4 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202205-01N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/20/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.2           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2021892-REC2 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9202201-11N  
Date Received: 02/19/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 02/19/92  
Analysis Date: 02/19/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.4           | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-KITREC4A SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205005-02N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92

Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
 Sample I.D.: 2043092-KITREC4B SCUDDER  
 Sample Type: Wipe

CHAS Lab #: 9205005-03N  
 Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92

Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
 Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-STOR216 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205004-06N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92  
Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-STORREC4 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205004-09N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92

Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-JC231 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205004-03N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92

Analysis Date: 05/02/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.

Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-JCSRNO6 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205004-04N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92  
Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-JC216 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205004-05N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92  
Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.

RCV BY:XEROX TELECOPIER 7010 ; 5- 5-92 5:43PM ;

6178481955-

SENT BY:CH Analyst Serv/Wood Rd; 5- 5-92 5:43PM ;

6178481955-



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-JCST2 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 92  
Date Received:

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92

Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units       |
|--------------------|-----|---------------|-------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq c |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq c |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq c |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq c |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq c |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq c |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq c |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-JCREG4 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205005-04N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92  
Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-JGREC4A SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205005-09N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92  
Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-JCSRNO5 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205005-05N  
Date Received: 05/01/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92  
Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2043092-JC201 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 920506  
Data Received: 05/03/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/01/92  
Analysis Date: 05/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 205692 L2A SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205064-04N  
Date Received: 05/07/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/07/92  
Analysis Date: 05/07/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 205692 L2B SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205064-05N  
Date Received: 05/07/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/07/92  
Analysis Date: 05/07/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 205692 L2C SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9205064-06N  
Date Received: 05/07/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 05/07/92  
Analysis Date: 05/07/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.1           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2-1, SCUDDER HALL  
Sample Type: Wipe

CHAS Lab #: 9112302-07N  
Date Received: 12/31/91

Polychlorinated Biphenyls (PCBs)

Extraction Date: 12/31/91  
Analysis Date: 12/31/91

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 1.0 | 20            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 100 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: SWS-10, SCUDDER HALL  
Sample Type: Wipe

CHAS Lab #: 9201009-10N  
Date Received: 01/02/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 01/02/91  
Analysis Date: 01/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 1.0 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 100 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: SWS-11, SCUDDER HALL  
Sample Type: Wipe

CHAS Lab #: 9201009-11N  
Date Received: 01/02/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 01/02/91  
Analysis Date: 01/03/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 1.0 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 100 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 202C1692, SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9201072-02N  
Date Received: 01/07/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 01/07/92  
Analysis Date: 01/07/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 1.0 | 15            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 203A1692 SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9201060-03N  
Date Received: 01/07/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 01/07/92  
Analysis Date: 01/07/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 1.0 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 1.0 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2042192-HS225F SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9204347-09N  
Date Received: 04/22/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/23/92

Analysis Date: 04/24/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2042292-HS2W SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9204370-07N  
Date Received: 04/23/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/25/92

Analysis Date: 04/27/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2042192-2N209W SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9204344-03N  
Date Received: 04/22/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/23/92  
Analysis Date: 04/23/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2042192-HN205F SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9204347-02N  
Date Received: 04/22/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/23/92

Analysis Date: 04/24/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2042292-HN2W SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9204370-03N  
Date Received: 04/23/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/25/92  
Analysis Date: 04/25/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.

Report: P C B P R E

Initials: EX Y/A

Status: D R A F T

Printed: APR. 29 1992 12:41 PM

Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2042192-2SS SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9204346-04N  
Date Received: 04/22/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/23/92  
Analysis Date: 04/24/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.

----- NOTES -----

Calculated and reported by KW.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2042192-2SC SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9204347-05N  
Date Received: 04/22/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/23/92

Analysis Date: 04/24/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | 0.2           | ug/100 sq cm |

Notes: ND = Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.



Client: Clean Harbors of Kingston, Inc.  
Sample I.D.: 2042192-2SN SCUDDER  
Sample Type: Wipe

CHAS Lab #: 9204347-10N  
Date Received: 04/22/92

Polychlorinated Biphenyls (PCBs)

Extraction Date: 04/23/92  
Analysis Date: 04/24/92

| Parameter          | MDL | Concentration | Units        |
|--------------------|-----|---------------|--------------|
| PCB - Aroclor 1016 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1221 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1232 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1242 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1248 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1254 | 0.1 | ND            | ug/100 sq cm |
| PCB - Aroclor 1260 | 0.1 | ND            | ug/100 sq cm |

Notes: ND - Below minimum detectable level (MDL)  
Wipe Area: 900 sq cm

The sample was mixed with hexane for 5 minutes. The resulting extract was analyzed by GC/ECD following EPA Method 8080.